

Phosphorus, Reactive (Orthophosphate) and Total

DOC316.53.01126

Ascorbic Acid Method

Method 10209/10210

6 to 60 mg/L PO₄³⁻ or 2 to 20 mg/L PO₄³⁻—P (UHR)

TNTplus 845

Scope and application: For wastewater, drinking water, boiler water, surface water and process analysis.



Test preparation

Instrument-specific information

Table 1 shows all of the instruments that have the program for this test. The table also shows the adapter and light shield requirements for the applicable instruments that can use TNTplus vials.

To use the table, select an instrument, then read across to find the applicable information for this test.

Table 1 Instrument-specific information for TNTplus vials

Instrument	Adapters	Light shield
DR6000, DR5000	—	—
DR3900	—	LZV849
DR3800, DR2800	—	LZV646
DR1900	9609900 or 9609800 (A)	—

Before starting

DR3900, DR3800, DR2800: Install the light shield in Cell Compartment #2 before this test is started.

Review the safety information and the expiration date on the package.

The recommended sample pH is 6–8.

The recommended temperature for samples and reagents is 15–25 °C (59–77 °F).

An incorrect result can occur if the sample and reagent are not at the recommended temperature.

The recommended temperature for reagent storage is 15–25 °C (59–77 °F).

The reagents that are used in this test contain molybdenum and are corrosive. Collect the reacted samples for proper disposal.

Use the DRB reactor with 13-mm wells for the digestion. If the reactor has 16-mm wells, put adapter sleeves into the wells.

DR1900: Go to All Programs>LCK or TNTplus Methods>Options to select the TNTplus number for the test. Other instruments automatically select the method from the barcode on the vial.

ATTENTION—Important information for the evaluation!

Without hydrolysis, only the (dissolved) orthophosphate is measured. The result of the orthophosphate measurement can be expressed as: mg/L PO₄-P (for example, process analysis), mg/L PO₄ (for example, drinking water or boiler water analysis), mg/L P₂O₅ (for example, soil analysis).

With hydrolysis, all of the phosphorus (Total-P, P_{total}) is measured. The result of the total phosphorus measurement can be expressed as: mg/L P_{tot} = Display mg/L PO₄-P (for example, for monitoring threshold values in wastewater), mg/L PO₄ (for example, drinking water or boiler water analysis), mg/L P₂O₅ (for example, soil analysis).

Inverting the vial after hydrolysis improves the reliability of the result.

Determination of orthophosphate: Filter the sample before the analysis.

Review the Safety Data Sheets (MSDS/SDS) for the chemicals that are used. Use the recommended personal protective equipment.

Dispose of reacted solutions according to local, state and federal regulations. Refer to the Safety Data Sheets for disposal information for unused reagents. Refer to the environmental, health and safety staff for your facility and/or local regulatory agencies for further disposal information.

Items to collect


Description	Quantity
Phosphorus, Reactive and Total UHR TNTplus Reagent Set	1
DRB200 reactor with 13-mm wells	1
Pipet, adjustable volume, 0.1–1.0 mL	1
Pipet tips, for 0.1–1.0 mL pipet	1
Test tube rack	1

Refer to [Consumables and replacement items](#) on page 6 for order information.


Sample collection and storage

- Collect samples in clean glass or plastic bottles that have been cleaned with 6 N (1:1) hydrochloric acid and rinsed with deionized water.
- Analyze the samples as soon as possible for best results.
- Do not use a detergent that contains phosphate to clean the sample bottles. The phosphate in the detergent will contaminate the sample.
- To preserve samples for later analysis, adjust the sample pH to 2 or less with concentrated sulfuric acid (approximately 2 mL per liter). Do not acidify samples to be analyzed only for reactive phosphorus. No acid addition is necessary if the sample is tested immediately.
- Keep the preserved samples at or below 6 °C (43 °F) for a maximum of 28 days (reactive phosphorus only: 48 hours).
- Let the sample temperature increase to room temperature before analysis.
- Before analysis, adjust the pH to 7 with 5 N sodium hydroxide solution.
- Correct the test result for the dilution caused by the volume additions.

Test procedure—total phosphorus




1. Set the DRB200 reactor power to on. Set the temperature to 100 °C (212 °F) or to 120 °C (248 °F).




2. Carefully remove the lid from the DosiCap™ Zip cap. Remove the cap from the test vial.




3. Use a pipet to add 0.4 mL of sample to the test vial.




4. Turn the DosiCap Zip over so that the reagent side goes on the test vial. Tighten the cap on the vial.




5. Shake the vial 2–3 times to dissolve the reagent in the cap. Look through the open end of the DosiCap to make sure that the reagent has dissolved.




6. Insert the vial in the preheated DRB200 reactor. Close the lid.




7. Keep the vial in the reactor for 1 hour at 100 °C (212 °F) or for 30 minutes at 120 °C (248 °F).




8. When the timer expires, carefully remove the vial from the reactor. Set the vial in a test tube rack. Let the temperature of the vial decrease to room temperature.




9. Shake the vial 2–3 times.



10. Use a pipet to add 0.5 mL of Solution B to the test vial. Immediately tighten the cap on the Solution B container.




11. Put a grey DosiCap C on the vial.




12. Tighten the cap on the vial and invert the vial 2–3 times.




13. Start the reaction time of 10 minutes.




14. When the timer expires, invert the vial 2–3 times.



15. Clean the vial.




16. DR1900 only: Select program 845. Refer to [Before starting](#) on page 1.




17. Insert the vial into the cell holder. DR1900 only: Push **READ**. Results show in mg/L PO_4^{3-} .


Test procedure—reactive phosphorus




1. Use a pipet to add 0.4 mL of sample to the test vial.



2. Use a pipet to add 0.5 mL of Solution B to the test vial. Immediately tighten the cap on the Solution B container.




3. Put a grey DosiCap C on the vial.




4. Tighten the cap on the vial and invert the vial 2–3 times.




5. Start the reaction time of 10 minutes.




6. When the timer expires, invert the vial 2–3 times.



7. Clean the vial.



8. DR1900 only: Select program 845. Refer to [Before starting](#) on page 1.



9. Insert the vial into the cell holder. DR1900 only: Push

READ.

Results show in mg/L PO₄³⁻.

Reagent blank correction

For the best results, measure the reagent blank value for each new lot of reagent. Replace the sample with deionized water in the test procedure to determine the reagent blank value. Subtract the reagent blank value from the sample results automatically with the reagent blank adjust option. Measure the reagent blank value when a new lot of reagent is used.

1. Use deionized water as the sample in the test procedure to measure the reagent blank value.
2. Set the reagent blank function to on. The measured reagent blank value is shown.
3. Accept the blank value. The reagent blank value is then subtracted from all results until the reagent blank function is set to off or a different method is selected.

Note: As an alternative, record or enter the reagent blank value at a different time. Push the highlighted reagent blank box and use the keypad to enter the value.

Sample blanks

Samples with color or turbidity can cause high results. Samples without color or turbidity do not require sample blanks. The digestion in the total phosphate test procedure usually removes all color and turbidity. A sample blank is not required. To adjust for color or turbidity in the reactive phosphate test procedure, use the steps that follow to find the sample blank.

1. Do the test procedure, but do not add the DosiCap C.
2. Put the cap on the vial, but do not remove the foil. Use the side of the cap that does not have the reagent.
3. Subtract the value from the final procedure step from the initial sample value to get the corrected sample concentration.

Note: Alternatively, samples that contain only turbidity can be filtered through a membrane filter, then analyzed.

Interferences

Table 2 shows that the ions were individually examined to the given concentrations and do not cause interference. No cumulative effects or influences of other ions were found. Verify the measurement results with sample dilutions or standard additions.

Table 2 Interfering substances

Interfering substance	Interference level
SO ₄ ²⁻	20,000 mg/L
Cl ⁻	10,000 mg/L
K ⁺ , Na ⁺ , Ca ²⁺	1000 mg/L

Table 2 Interfering substances (continued)

Interfering substance	Interference level
Mg ²⁺ , NO ₃ ⁻	500 mg/L
Co ²⁺ , Fe ²⁺ , Fe ³⁺ , Zn ²⁺ , Cu ²⁺ , Ni ²⁺ , I ⁻ , NO ₂ ⁻ , Cd ²⁺ , Sn ⁴⁺ , NH ₄ ⁺ , Mn ²⁺ , Al ³⁺ , Hg ²⁺ , Pb ²⁺ , SiO ₂	50 mg/L
Ag ⁺	25 mg/L
Cr ³⁺	10 mg/L
Cr ⁶⁺	5 mg/L

Accuracy check

Standard solution method

Use the standard solution method to validate the test procedure, the reagents and the instrument.

Items to collect:

- Phosphate Standard Solution, 50-mg/L PO₄³⁻ or Wastewater Influent Standard Solution, Mixed Parameter (contains 10-mg/L PO₄³⁻)
- Use the test procedure to measure the concentration of the standard solution.
 - Compare the expected result to the actual result.

Method performance

The method performance data that follows was derived from laboratory tests that were measured on a spectrophotometer during ideal test conditions. Users can get different results under different test conditions.

Program	Standard	Precision (95% confidence interval)	Sensitivity Concentration change per 0.010 Abs change
barcode	50.0 mg/L PO ₄ ³⁻	49–51 mg/L PO ₄ ³⁻	—

Summary of Method

Phosphates present in organic and condensed inorganic forms (meta-, pyro- or other polyphosphates) are first converted to reactive orthophosphate in the total phosphorus procedure. Treatment of the sample with acid and heat provides the conditions for hydrolysis of the condensed inorganic forms. Organic phosphates are also converted to orthophosphates in the total phosphorus procedure by heating with acid and persulfate. The reactive phosphorus procedure measures only the reactive (ortho) phosphorus present in the sample. The reactive or orthophosphate ions react with molybdate and antimony ions in an acidic solution to form an antimonyl phosphomolybdate complex, which is reduced by ascorbic acid to phosphomolybdenum blue. The measurement wavelength is 880 nm (DR 1900: 714 nm).

Consumables and replacement items

Required reagents

Description	Quantity/Test	Unit	Item no.
Phosphorus, Reactive and Total UHR TNTplus Reagent Set	1	25/pkg	TNT845

Required apparatus

Description	Quantity/test	Unit	Item no.
DRB200 Reactor, 115 VAC option, 9 x 13 mm + 2 x 20 mm, 1 block	1	each	DRB200-01
DRB200 Reactor, 230 VAC option, 9 x 13 mm + 2 x 20 mm, 1 block	1	each	DRB200-05

Required apparatus (continued)

Description	Quantity/test	Unit	Item no.
Pipet, adjustable volume, 0.1–1.0 mL	1	each	BBP078
Pipet tips, for 0.1–1.0 mL pipet	2	100/pkg	BBP079
Test tube rack	1	each	1864100
Light shield, DR3800, DR2800, DR2700	1	each	LZV646
Light shield, DR3900	1	each	LZV849

Recommended standards

Description	Unit	Item no.
Phosphate Standard Solution, 50-mg/L as PO_4^{3-}	500 mL	17149
Wastewater Influent Standard Solution, Mixed Parameter, for $\text{NH}_3\text{-N}$, $\text{NO}_3\text{-N}$, PO_4^{3-} , COD, SO_4^{2-} , TOC	500 mL	2833149

Optional reagents and apparatus

Description	Unit	Item no.
Filter membrane, 0.45 micron, 25 mm	100/pkg	2514101
Flask, volumetric, Class A, 1000 mL glass	each	1457453
Hydrochloric Acid Solution, 6.0 N (1:1)	500 mL	88449
Reactor adapter sleeves, 16 mm to 13 mm diameter, for TNTplus vials	5/pkg	2895805
Sampling bottle with cap, low density polyethylene, 500 mL	12/pkg	2087079
Sodium Hydroxide Standard Solution, 5.0 N	100 mL MDB	245032
Sulfuric Acid, concentrated, ACS	500 mL	97949



FOR TECHNICAL ASSISTANCE, PRICE INFORMATION AND ORDERING:

In the U.S.A. – Call toll-free 800-227-4224

Outside the U.S.A. – Contact the HACH office or distributor serving you.

On the Worldwide Web – www.hach.com; E-mail – techhelp@hach.com

HACH COMPANY
WORLD HEADQUARTERS
Telephone: (970) 669-3050
FAX: (970) 669-2932