



**STABLCAL<sup>®</sup> FORMAZIN REFERENCE SUSPENSIONS\***  
**CERTIFICATE OF COMPLIANCE**  
**StablCal<sup>®</sup> Calibration Set # 2897200**

TEST PERFORMED: The Standards listed below were measured using a Hach ISO Turbidimeter. This instrument was calibrated on fresh formazin standards that are traceable to the Stock Primary Standard Formazin Suspension as specified in Standard Methods, Section 2130 B, part 3b.

	<b>Cat. No.</b>	<b>Value</b>	<b>Variation</b>
Standard <0.1	2659742	<0.1 NTU	n/a
Standard I	2896742	3 NTU	+/- 0.15 NTU
Standard II	2896842	6 NTU	+/- 0.30 NTU
Standard III	2896942	18 NTU	+/- 0.90 NTU
Standard IV	2897042	30 NTU	+/- 1.50 NTU

\*StablCal<sup>®</sup> Stabilized Formazin Standards are cited as primary calibration standards in Hach Method 8195, an acceptable version of USEPA Method 180.1. StablCal Formazin Reference Suspensions I to IV according to Pharm Eur. 5.0/2.2.1.

Download Certificates of Analysis by entering the StablCal Calibration Set part number and lot number on our web site

Signed:

on behalf of Hach Company

Hach Cat. No. 2897287