

## PART 1 GENERAL

- 1.1 Section includes:
  - A. Instrument for monitoring turbidity in water accordance/compliance with DIN EN ISO 7027.
- 1.2 Measurement Procedures
  - A. The method of measuring turbidity will be nephelometric using pulse scattered infrared light at 860nm at a 90° angle in accordance with/compliance with DIN EN ISO 7027.
- 1.3 Alternates
  - A. Other methods of turbidity measurement including those that require a sample cell, those with incandescent light sources, or turbidimeters used for EPA reporting are not acceptable.
- 1.4 System Description
  - A. Performance Requirements
    1. Range
      - a. 0.0001 to 1000 FNU (1 FNU = 1 NTU)
    2. Resolution
      - a. 0.0001 to 0.9999 / 1.000 to 9.999 / 10.00 to 99.99 / 100 to 1000 FNU (NTU)
    3. Precision
      - a.  $\pm 0.008$  FNU or  $\pm 1\%$  of reading (0 to 10 FNU)
    4. Repeatability
      - a.  $\pm 0.003$  FNU or  $\pm 5\%$  of reading (0 to 2 FNU)
    5. Response time
      - a. 1 to 60 seconds (user adjustable)
- 1.5 Certifications
  - A. EMC: CE compliant for conducted and radiated emissions CISPR 11 (Class A limits), EMC Immunity EN 61326-1 (Industrial limits) when connected to an sc controller.
  - B. Safety: General Purpose UL/CSA 61010-1 with cETLus safety mark when connected to an sc controller
  - C. Australian C-TICK and Korean KC Markings when connected to an sc controller.
  - D. IP 55 Enclosure Rating
- 1.6 Environmental Requirements
  - A. Operational Criteria
    1. Operating Temperature
      - a. 36 to 104 °F (2 to 40°C)
    2. Sample Temperature
      - a. 122 °F (50 °C) maximum
    3. Sample Pressure
      - a. 87 psi at 68°F (6 bar at 20°C)
    4. Sample flow rate
      - a. Minimum: 0.2 L/min
      - b. Maximum: 1L/min
    5. Sample Salt Content (for *seawater* version ONLY)
      - a. Tested up to 65 g/L

- 1.7 Warranty
  - A. The sensor includes a one-year warranty from the date of shipment.
  
- 1.8 Maintenance Service
  - A. Scheduled Maintenance
    - 1. Every 1200 Cycles
      - a. Replace wiper profile (only on *plus* and *seawater* versions)
    - 2. Every Two Years
      - a. Replace desiccant
  - B. Unscheduled Maintenance
    - 1. Clean measuring chamber
      - a. Dependent on substances contained in the water
    - 2. Check Zero Point
      - a. Dependent on substances contained in the water
    - 3. Check Gradient
      - a. At least once per year

## PART 2 PRODUCTS

- 2.1 Manufacturer
  - A. Hach-Lange GmbH, Berlin, Germany
    - 1. Ultraturb sc Basic/Plus/Seawater Turbidimeter
  
- 2.2 Manufactured Unit
  - A. The Ultraturb sc Turbidimeter consists of an 860nm LED light source, detection system, and internal light trap. Sample chamber wiper available for *plus* and *seawater* versions.
  
- 2.3 Equipment
  - A. The Ultraturb sc functions when attached to Hach model sc200 or sc1000 controllers only. (Additional specifications can be found in the CSI documents for these particular controllers)
  - B. The Ultraturb sc turbidimeter operates continuously.
  - C. The Ultraturb sc turbidimeter provides user selectable bubble rejection, alarm and controller output hold, and self-test diagnostics.
  - D. The sc200 controller is capable of functioning with one or two Ultraturb sc turbidimeters; the sc1000 controller is capable of functioning with up to eight Ultraturb sc turbidimeters.
  - E. Wetted materials as follows:
    - 1. Measuring window:
      - a. Quartz
    - 2. Measuring Chamber:
      - a. Noryl GFN2
    - 3. Wiper axle:
      - a. Stainless Steel 1.4571
    - 4. Wiper arm (*seawater* version only):
      - a. Titanium Alloy
    - 5. Wiper profile
      - a. Silicone

- 2.4 Components
  - A. Standard Equipment
    - 1. Ultraturb sc sensor with appropriate cable length
    - 2. User Manual
    - 3. Factory Test Certificate
    - 4. Accessory Set
    - 5. Wiper Set (only for *plus* and *seawater* versions)
  - B. Dimensions: 9.9 x 9.4 x 4.3 in. (250 x 240 x 110 mm)
  - C. Weight: 3.3 lbs (1.5 kg)

- 2.5 Optional Accessories
  - A. Extension Cable
  - B. Filters for Zero Point Calibration
  - C. Formazin Turbidity Standard

### PART 3 EXECUTION

#### 3.1 Preparation

- 1. Mounting
  - a. Wall mount only
- 2. Sample inlet
  - a. 13 mm ID tubing
- 3. Drain
  - a. 13 mm ID tubing

#### 3.2 Installation

- A. Contractor will install the analyzer in strict accordance with the manufacturer's instructions and recommendation.
- B. Manufacturer's representative will include a half-day of start-up service by a factory-trained technician, if requested.
  - 1. Contractor will schedule a date and time for start-up.
  - 2. Contractor will require the following people to be present during the start-up procedure.
    - a. General contractor
    - b. Electrical contractor
    - c. Hach Company factory trained representative
    - d. Owner's personnel
    - e. Engineer

#### 3.3 Manufacturer's Service and Start-Up

- A. Contractor will include the manufacturer's services to perform start-up on instrument to include basic operational training and certification of performance of the instrument.

- B. Contractor will include a manufacturer's Service Agreement that covers all the manufacturer's recommended preventative maintenance, regularly scheduled calibration and any necessary repairs beginning from the time of equipment startup through to end user acceptance / plant turnover and the first 12 months of end-user operation post turnover.
- C. Items A and B are to be performed by manufacturer's factory-trained service personnel. Field service and factory repair by personnel not employed by the manufacturer is not allowed.
- D. Use of manufacturer's service parts and reagents is required. Third-party parts and reagents are not approved for use.

END OF SECTION