APC341
Nitrite
MR 0,015 – 0,6 mg/L NO_2-N / 0,05 – 2,0 mg/L NO_2

DOC312.53.94342

Principle
Nitrites react with primary aromatic amines in acidic solution to form diazonium salts. These combine with aromatic compounds that contain an amino group or a hydroxyl group to form intensively colored azo dyes.

Range of Application
Waste water, drinking water, table water, surface water, mineral water

Interferences

T1


2000 mg/l:	Cl^-, SO_4^{2-}
1000 mg/l:	K^+, NO_3^-
500 mg/l:	NH_4^+, PO_4^{3-}, Ca^{2+}
100 mg/l:	Mg^{2+}
50 mg/l:	Cr^{3+}
25 mg/l:	Co^{2+}, Zn^{2+}, Cd^{2+}, Mn^{2+}, Hg^{2+}
12 mg/l:	Ni^{2+}
10 mg/l:	Ag^+, Fe^{2+}
5 mg/l:	Sn^{4+}, Fe^{3+}


The ions listed in T1 have been individually checked up to the given concentrations and do not cause interference. We have not determined cumulative effects and the influence of other ions.

Chromium(VI) ions interfere with the determination. Copper(II) ions interfere with the determination even at concentrations below 1 mg/L.
The measurement results must be subjected to plausibility checks (dilute and/or spike the sample).

Sample Volume	**2 mL**
Temperature Sample/sample cuvette	**15 – 25°C**
pH sample	**3 - 10**

Method Library:
APC341 is pre-programmed in the method library. Please check under Settings/Software/Application/Methods **Nitrite** and Tests **APC341**.





Run the APC341 Nitrite method

Create a Run like described in the QUICK GUIDE

- Place the APC341 cuvettes according to the settings in the Software in the cuvette racks.
- Place the samples according to the settings in the Software in the sample racks
- Check if enough Rinsing/Dilution water is available
- Initialize the AP 3900 multi and the Dispenser

HACH LANGE GMBH
Willstätterstraße 11
D-40549 Düsseldorf
Tel. +49 (0) 2 11 52 88-0
Fax +49 (0) 2 11 52 88-143
info-de@hach.com
www.hach.com

© Hach Company, 2018. All rights reserved.