



**STABLCAL® STABILIZED FORMAZIN STANDARDS\***  
**CERTIFICATE OF COMPLIANCE**  
**StablCal® Calibration Set # 2971205**

TEST PERFORMED: The Standards listed below were measured using a Hach 2100 Portable Turbidimeter. The instrument was calibrated on fresh formazin standards that are traceable to the Stock Primary Standard Formazin Suspension as specified in Standard Methods, Section 2130 B, part 3b.

|             | <b>Cat. No.</b> | <b>Value</b> | <b>Variation</b> |
|-------------|-----------------|--------------|------------------|
| Standard #1 | 2961801         | 10 NTU       | +/- 0.5 NTU      |
| Standard #2 | 2684801         | 20 NTU       | +/- 1.0 NTU      |
| Standard #3 | 2684901         | 100 NTU      | +/- 5.0 NTU      |
| Standard #4 | 2660501         | 800 NTU      | +/- 40 NTU       |

\*StablCal® Stabilized Formazin Standards are cited as primary calibration standards in Hach Method 8195, an acceptable version of USEPA Method 180.1.

Download Certificates of Analysis by entering the StablCal Calibration Set part number and lot number on our web site.

Signed: \_\_\_\_\_

  
on behalf of Hach Company

Hach Cat. No. 2971287