LCK 337 Nickel

DOC312.53.94049

0.1–6.0 mg/L Ni or 0.12–7.20 mg/L Ni (Crack-Set LCW 902) LCK 337

Scope and application: For wastewater, drinking water and process analysis.

Test preparation

Test storage

Storage temperature: 15–25 °C (59–77 °F)

pH/Temperature


The pH of the water sample must be between pH 3–10.

The temperature of the water sample and reagents must be between 15–25 °C (59–77 °F).


Before starting

Undissolved nickel or nickel contained in complexes can only be determined after digestion with Crack-Set LCW 902.
Review safety information and expiration date on the package.
Review the Safety Data Sheets (MSDS/SDS) for the chemicals that are used. Use the recommended personal protective equipment.
Dispose of reacted solutions according to local, state and federal regulations. Refer to the Safety Data Sheets for disposal information for unused reagents. Refer to the environmental, health and safety staff for your facility and/or local regulatory agencies for further disposal information.


Procedure




1. Carefully pipet **2.0 mL** of **sample**.




2. Close the cuvette and invert a few times until the freeze-dried contents are **completely dissolved**.




3. Open the cuvette.




4. Carefully pipet **0.2 mL** of **solution A**.



5. Close the cuvette and invert a few times.



6. After **3 minutes**, thoroughly clean the outside of the cuvette and evaluate.



7. Insert the cuvette into the cell holder.
DR 1900: Go to LCK/TNTplus methods.
Select the test, push **READ**.

Interferences

The ions listed in the table have been individually checked against the given concentrations and do not cause interference. The cumulative effects and the influence of other ions have not been determined.

The measurement results must be subjected to plausibility checks (dilute and/or spike the sample).

Removal of Interferences

Nickel concentrations greater than the measuring range cause precipitation in the cuvette. In such cases the water sample must first be diluted with distilled water.

Interference level	Interfering substance
1000 mg/L	K^+, Na^+, Cl^-, SO_4^{2-}
500 mg/L	NH_4^+, NO_3^-, Ca^{2+}, PO_4^{3-}, CO_3^{2-}
50 mg/L	Cr^{6+}, Zn^{2+}, F^-, NO_2^-
10 mg/L	Al^{3+}, Cr^{3+}, Cd^{2+}, Co^{2+}, Sn^{2+}, Pb^{2+}
5 mg/L	Fe^{2+}, Fe^{3+}, Mn^{2+}, Cu^{2+}, Mg^{2+}, Hg^{2+}
1 mg/L	Ag^+

Summary of method

In the presence of an oxidizing agent, nickel ions react with dimethylglyoxime in an alkaline solution to form an orange-brown-colored complex.

HACH LANGE GMBH
Willstätterstraße 11
D-40549 Düsseldorf
Tel. +49 (0) 2 11 52 88-0
Fax +49 (0) 2 11 52 88-143
info-de@hach.com
www.hach.com

© Hach Company, 2019. All rights reserved.
03/2019, Edition 1