Quality certificate

Technical data for cuvette test APC342
(Results as NO_2)

Sensitivity	0.056 Abs./(mg/L)
Ordinate intersect	0.089 Abs.
Residual standard deviation	0.0061 Abs.
Method variation coefficient	1.00 %
Method standard deviation	0.108 mg/L
Confidence interval (95%)	± 0.26 mg/L


The technical data for cuvette test APC342 were determined in conformity with ISO 8466-1 and DIN 38402 A51 „Calibration of analysis methods“.

The series of the smallest and largest calibration standards exhibit normal distribution and are outlier- and trend-free.
The calibration gives a linear function.

Technical data in conformity with DIN 32645

Detection limit	0.08 mg/L
Quantitation limit	0.24 mg/L

The detection and the quantitation limits were determined in conformity with DIN 32645.



Result	Confidence intervall (95%)
4.0 mg/L	± 0.28 mg/L
8.0 mg/L	± 0.26 mg/L
12.0 mg/L	± 0.26 mg/L
16.0 mg/L	± 0.27 mg/L
20.0 mg/L	± 0.29 mg/L

HACH LANGE GmbH
BU Lab Application

Dr. Ralf Kloos

UNITED FOR WATER QUALITY

HACH LANGE GmbH
Willstätterstraße 11
D-40549 Düsseldorf
Tel.+49 (0)211-52 88-320
Fax+49 (0)211-52 88-143
Kundenservice@hach-lange.de
www.hach-lange.de