



STABLCAL® STABILIZED FORMAZIN STANDARDS*
CERTIFICATE OF COMPLIANCE
StablCal® Calibration Set # 2659505

TEST PERFORMED: The Standards listed below were measured using a Hach EPA Turbidimeter. This instrument was calibrated on fresh formazin standards that are traceable to the Stock Primary Standard Formazin Suspension as specified in Standard Methods, Section 2130 B, part 3b.

	Cat. No.	Value	Variation
Standard #1	2659701	<0.1 NTU	n/a
Standard #2	2660101	20 NTU	+/- 1.0 NTU
Standard #3	2660401	200 NTU	+/- 10 NTU
Standard #4	2660601	1000 NTU	+/- 50 NTU
Standard #5	246102	4000 NTU	+/- 200 NTU
Standard #6	2584201	7500 NTU	+/- 375 NTU

*StablCal® Stabilized Formazin Standards are cited as primary calibration standards in Hach Method 8195, an acceptable version of USEPA Method 180.1.

Download Certificates of Analysis by entering the StablCal Calibration Set part number and lot number on our web site.

Signed: 
on behalf of Hach Company

Hach Cat. No. 2659587